PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

SWSCALE

2320 East Jones Ave, Phoenix, AZ 85040

and hereby declares that the Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Mass, Force, and Weighing Devices Calibration

(As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Initial Accreditation Date:	*Issue Date:*	*Expiration Date:*
September 07, 2023	June 06, 2026	June 06, 2028

Tracy Szerszen
President

Accreditation No.:	*Certificate No.:*
122385	L26-494-1

Perry Johnson Laboratory Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlabs.com

Certificate of Accreditation: Supplement

SWSCALE

2320 East Jones Ave, Phoenix, AZ 85040
Contact Name: Timothy Johnson Phone: 602-550-0481

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Mass, Force, and Weighing Devices	Class I Balances	Up to 300 g (Res = 0.000 1 g)	0.64 mg	Class 1 Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class II Balances	Up to 500 g (Res =0 .001 g)	6 mg	Class F Weights Class 5 Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class II Balances	Up to 3 kg (Res =0.002 g)	27 mg	Class F Weights Class 5 Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class II Balances	Up to 6 kg (Res = 0.005 g)	54 mg	Class F Weights Class 5 Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 150 kg (Res = 0.02 kg)	0.06 kg	Class F Weights Class 5 Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 1 500 kg (Res = 0.1 kg)	0.13 kg	Class F Weights Class 5 Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 2 500 kg (Res = 0.2 kg)	0.26 kg	Class F Weights Class 5 Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 20 lb (Res = 0.002 lb)	0.004 lb	Class F Weights Class 5 Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 100 lb (Res = 0.01 lb)	0.01 lb	Class F Weights Class 5 Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 200 lb (Res = 0.02 lb)	0.02 lb	Class F Weights Class 5 Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M

Certificate of Accreditation: Supplement

SWSCALE

2320 East Jones Ave, Phoenix, AZ 85040
Contact Name: Timothy Johnson Phone: 602-550-0481

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Mass, Force, and Weighing Devices	Class III Scales	Up to 1 000 lb (Res = 0 .1 lb)	0.13 lb	Class F Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 2 000 lb (Res = 0.2 lb)	0.27 lb	Class F Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 5 000 lb (Res = 0.5 lb)	0.67 lb	Class F Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 10 000 lb (Res = 1 lb)	1.3 lb	Class F Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 20 000 lb (Res = 2 lb)	2.6 lb	Class F Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 50 000 lb (Res = 5 lb)	6.5 lb	Class F Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 100 000 lb) (Res = 10 lb)	12 lb	Class F Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III Scales	Up to 200 000 lb (Res = 20 lb)	26 lb	Class F Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III L Scales	Up to 200 000 lb (Res = 20 lb)	26 lb	Class F Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M
Mass, Force, and Weighing Devices	Class III L Scales	Up to 400 000 lb (Res = 50 lb)	58 lb	Class F Weights	NIST Handbook 44 NISTR 6919 EURAMET cg-18	F1, F2	F, O, M

Certificate of Accreditation: Supplement

SWSCALE

2320 East Jones Ave, Phoenix, AZ 85040

Contact Name: Timothy Johnson Phone: 602-550-0481

Accreditation is granted to the facility to perform the following conformity assessment activities:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.

2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.

3. Location of activity:

Location Code	**Location**
F	Conformity assessment activity is performed at the CABs fixed facility
O	Conformity assessment activity is performed onsite at the CABs customer location
M	Conformity assessment activity is performed from a mobile facility

4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.

5. This is the primary site for all quality management system activities.